

REV: 6/11

PRODUCT INFORMATION SUMMARY

Human IL-33 "SUPER X" ELISA Kit

Product Number **RHF632CKX**
5 Plate Kit (5 x 96 tests)
RHF632CKX2 (2 Plate Kit)
RHF632CKX1 (Single Plate)

Items Provided:

PRE-COATED ELISA Plates (1,2 or 5 plates)
Biotin-Labeled tracer 25.0 ug or 0.5 mL (**see vial**)
Antigen Standard **1.0 ug** (or 2.0 ug - see vial)
Streptavidin-HRP 0.5 mL
TMB Substrate 25 mL x 2 (15 mL - 2 plate)
Wash Buffer Concentrate 125 mL (20X) (50 mL - 2 plate)

DESCRIPTION:

Human Interleukin 33 (IL-33) **SUPER X ELISA Kit** provides one, two or five **PRE-COATED** microplates (coated with antigen affinity purified capture antibody stabilized with our proprietary ELISA coating/blocking reagent. A biotin labeled tracer antibody, antigen standard, HRP developing reagents and wash buffer are included.

Reactivity with various sample types, including serum/plasma samples, is evaluated by customer.

Note: Reconstitute components only when ready to run assay.

TRACER ANTIBODY:

Provided as 25 ug (lyophilized)or as **0.5 mL liquid @ 33.0 ug/mL (see vial)** of Biotin labeled, antigen-affinity purified antibody, additive-free. Reconstitute the 25 ug (lyophilized vial) in 500 uL sterile water **containing 0.1% BSA.** (FREEZE aliquots for long-term storage). -

****For liquid vial store refrigerated only (contains preservative)**.**

STANDARD: Provided as 1.0 ug or 2.0 ug (**see vial**) of recombinant Human IL-33 (see vial). Quick-spin and **reconstitute in distilled water** (pH 8.0) - concentration approx. 100.0 ug/mL. Further dilutions can be made in 0.05% Tween-20, 0.1% BSA in PBS.

DEVELOPING REAGENTS:

- Streptavidin-HRP (S100180CX) 1.0 mL - **store @ -20 Deg. C.**
- TMB Substrate Solutions - Part A and Part B (25.0 mL each)
cat # ES200CX
- Wash Buffer Concentrate (20X concentrate) Mix 1 volume of wash buffer with 19 volumes of distilled water. Stable for one month @ 4 Deg. C. once mixed to working volume.

HANDLING/ STORAGE: Reconstitute reagents when ready to build ELISA assay. Biotin Tracer Antibody provided as lyophilized, once reconstituted- can be stored for approximately one month at 4 Degrees C. Or store **frozen** at -20 Degrees C. for up to 6 months.

Biotin tracer antibody provided as 0.5 mL liquid - STORE refrigerated only - contains preservative.

Standard (rec. IL-33) can be stored in liquid state (@ 4 Deg. C.) For up to one week, or store **frozen, with addition of 0.1% BSA,** at -20 Deg. C. for up to 2 months. AVOID repeat freeze-thaw. Pre-coated ELISA plates should be stored refrigerated (@ 4 Deg.C.) in sealed plastic bag with desiccant pack, and are stable for one year from date of receipt. If needed, additional components (antigen standard or developing reagents) can be purchased as separate items to run additional stored plates - inquire.

Store Streptavidin-HRP (S100180CX) **frozen** @ -20 Deg. C.

Store TMB solutions and wash buffer concentrate at 4 Deg. C.
CAUTION: Substrate Solution B contains 20% acetone. FLAMMABLE.
Keep away from sources of heat or flame.

MATERIALS RECOMMENDED:

Tween -20.
BSA (ELISA grade only, protease-free)
Dubelco's PBS (10X)
PBS: Dilute to 1XPBS in sterile water
Diluent: use : 0.05% Tween-20, 0.1% BSA in PBS
2N Sulfuric acid - Stop Solution.

PLATE PREPARATION:

Plates are **pre-coated** with capture antibody and blocked/stabilized with ANTIGENIX proprietary ELISA coating stabilizer (EA150) and are **ready to use!**

Store plates refrigerated (@ 4 Deg. C.) in sealed plastic bags with desiccant pack. Plates can be stored for one year from data of receipt.

PROTOCOL:

STANDARD/SAMPLE: Dilute a portion of the standard (store unused standard in aliquots, high concentration, frozen -20 Deg. C. with addition of 0.1% BSA) from **4.0 ng/mL** to zero in diluent (serial dilution). Immediately add 100 uL of standard or sample to each well in duplicate. Incubate at room temp. for **1 hour**.

DETECTION: Aspirate and wash plate 4 times. **Dilute** portion of detection (Biotin Tracer) antibody in diluent to concentration of **0.20 ug/mL**. Add 100 uL per well. **Incubate** at room temperature for **1 hour**.

STREPTAVIDIN-HRP: Aspirate and wash plate 4 times. **Dilute** Streptavidin-HRP conjugate approx. **1:2,000** in diluent. Add 100 uL per well, incubate 30 minutes at room temperature.

SUBSTRATE: Prepare substrate solution no more than 15 minutes before last incubation of assay: Mix one part TMB Solution A with one part TMB Solution B in a clean container. If, upon mixing, TMB solution turns blue - TMB solution is **contaminated-DO NOT USE**. Use mixed substrate solution **WITHIN 2 Hours, and AVOID DIRECT LIGHT**.

Aspirate and wash plate 4 times. **Note:** Wash steps are critical! **Add** 100 uL **substrate solution** to each well. **Incubate** at room temp. for color development. **Add** 100 uL of **Stop solution** (2N Sulfuric Acid) **within 15-20 minutes** to stop color development - gently tap plate to mix. Read plate at 450 nm within 30 minutes of addition of stop solution.

NOTE: reliable standard curves are obtained when O.D. readings do not exceed 0.25 units for the zero standard concentration, or 1.6 units for the highest standard concentration. Monitor the plate every 5 minutes for approximately 30 minutes.

X-REACTIVITY DATA:

The following factors were tested at 40-50 ng/mL:

Minimal X-Reactivity was observed (2%) with Mouse IL-33.

No significant X-Reactivity was observed with:

Human: IL-13, IL-13 variant, IL-1 alpha and beta; IL-4; IL-5;
IFN Gamma.

RESEARCH USE ONLY -NOT For DIAGNOSTIC USE

ANTIGENIX AMERICA Inc.

WWW.ANTIGENIX.COM