



**Cytoskeleton, Inc.**

# Manual

V. 6.0

## Precision Red Advanced Protein Assay Reagent

**Cat. # ADV02**



# Manual Contents

---

<b>Section I: Introduction</b>	
Background .....	5
Uses .....	5
<b>Section II: Purchaser Notification</b> .....	6
<b>Section III: Kit Contents</b> .....	7
Contents .....	7
Equipment Required .....	7
<b>Section IV: Quick Methods</b> .....	8
A) Protein solutions between 0.3 to 10 mg/ml .....	8
B) Low concentration protein solutions (<0.3 mg/ml) .....	8
C) High Concentration protein solutions (10-100 mg/ml) .....	8
<b>Section V: Detailed Methods</b> .....	9
A) Description of the assay characteristics .....	9
B) Standard Assay .....	9
C) Method for measurements in 96-well format .....	10
<b>Section VI: Compatibility table</b> .....	11
<b>Section VII: Troubleshooting</b> .....	12
<b>Section VIII: Notes on Updated Version</b> .....	13



# I: Introduction

---

## **Background**

The Precision Red Advanced Protein Assay Reagent (Cat# ADV02) is designed to extend the linear range of protein measurement, a limiting factor in other protein assays. The reagent combines the useful properties of low protein to protein variance and a broad linear response to many types of purified protein. A simple one step procedure results in a red to purple/blue color change characterized by an increase in absorbance at 580 to 620 nm.

## **Uses**

- 1) To determine protein concentration in a variety of common buffers.
- 2) To determine protein concentration in biological fluids (e.g. serum or saliva).
- 3) To determine protein concentration in tissue culture media.
- 4) To follow protein purification steps in combination with activity assays.

## II: Purchaser Notification

---

### **Limited Use Statement**

The purchase of this product conveys to the buyer the non-transferable right to use the purchased amount of product and components of product in research conducted by the buyer. The buyer cannot sell or otherwise transfer this product or any component thereof to a third party or otherwise use this product or its components for commercial purposes. Commercial purposes include, but are not limited to: use of the product or its components in manufacturing; use of the product or its components to provide a service; resale of the product or its components.

The terms of this Limited Use Statement apply to all buyers including academic and for-profit entities. If the purchaser is not willing to accept the conditions of this Limited Use Statement, Cytoskeleton Inc. is willing to accept return of the unused product with a full refund.

### III. Kit Contents

---

**Contents:** 1 x 500 ml Advanced Protein Assay Reagent (1X concentrate; Cat # ADV02).  
Store at room temperature for up to 6 months.

**Equipment Required:**

- 1) Spectrophotometer capable of measuring 580 to 620 nm wavelength (optimally 600 nm)
- 2) Small volume cuvettes (1.0 ml) or 96-well plates.
- 3) Pipettors 3-20  $\mu$ l, 100-1000  $\mu$ l and 10 ml capacity.
- 4) 1.5 or 15 ml disposable tubes.

## VI: Quick Methods

---

### A) Protein solutions between 0.3 to 10 mg/ml

1. Pipette 1 ml of ADV02 into a 1.5 ml disposable microfuge tube.
2. Pipette 10  $\mu$ l of protein solution into the tube containing 1.0 ml ADV02 and mix by inverting.
3. Incubate at room temperature for one minute.
4. Blank the spectrophotometer on ADV02 at 600 nm and read absorbance of your sample.
5. Calculate protein concentration based on 1.00 OD<sub>600 nm</sub> = 100  $\mu$ g protein per ml reagent per cm, and multiply this protein concentration by 100 to achieve the protein concentration in  $\mu$ g/ml of original protein solution.

### B) Low concentration protein solutions (<0.3 mg/ml)

1. Pipette 900  $\mu$ l of ADV02 into a 1.5 ml disposable microfuge tube.
2. Pipette 100  $\mu$ l of protein solution into the tube containing 900  $\mu$ l ADV02 and mix by inverting.
3. Incubate at room temperature for one minute.
4. Blank the spectrophotometer on 1 ml of 900  $\mu$ l ADV02 plus 100  $\mu$ l of your buffer at 600 nm and read absorbance of your sample.
5. Calculate protein concentration based on 1.00 OD<sub>600 nm</sub> = 100  $\mu$ g protein per ml reagent per cm, and multiply this protein concentration by 10 to achieve the protein concentration in  $\mu$ g/ml of original protein solution.

### C) High concentration protein solutions (10-100 mg/ml):

1. Pipette 10 ml of ADV02 into a 15 ml disposable plastic tube.
2. Pipette 10  $\mu$ l of protein solution into the tube containing 10 ml ADV02 and mix thoroughly.
3. Incubate at room temperature for one minute.
4. Blank the spectrophotometer on ADV02 and read absorbance of your sample at 600 nm.
5. Calculate protein concentration based on 1.00 OD<sub>600 nm</sub> = 100  $\mu$ g protein per ml reagent per cm, and multiply this protein concentration by 1000 to achieve the protein concentration in  $\mu$ g/ml of original protein solution.

**Note:** For 96-well plates and 300  $\mu$ l wells, a 300  $\mu$ l volume is equivalent to 0.8 cm light pathlength. See below for the method suited to 96-well plate technology.



# VI: Detailed Methods

## A) Description of the assay characteristics

The linear range of the ADV02 assay is between 0.05 OD600 nm to 0.80 OD600 nm, either side of this range will result in inaccurate readings. The standard factor (1.00 OD600 nm is equivalent to 100 µg protein per ml of reagent) is used to calculate protein concentration based on the average value obtained from several different proteins. The highest value is found with BSA where an OD of 1.00 is equivalent to 90 µg/ml, and the lowest with IgG where an OD of 1.00 is equivalent to 130 µg/ml. Greater than 95% of tested proteins have been found to be located between these values.


## B) Standard Assay

The linear range for ADV02 is up to 100 µg protein per ml reagent. Therefore if your protein concentration exceeds this level in the final mixture then you will be underestimating its true concentration. The following formula is used as a general rule to estimate protein concentration:

$$1.00 \text{ OD } 600 \text{ nm} = 100 \text{ } \mu\text{g protein per ml reagent per cm light pathlength}$$

For 95% of individually purified proteins this equation will estimate its concentration to within 20% of the true value. For protein extracts virtually all samples will be estimated to within 20% of the true value.

For a more accurate determination, especially at low protein concentrations, it is recommended to use a standard curve of IgG protein, as shown in Figure 1.



**Figure 1: Standard curves of Actin, BSA, IgG and Tubulin protein using the Precision Red Advanced Protein Assay Reagent**

## V: Detailed Methods (continued)

---

### C) Method for measurements in 96-well format:

1. Pipette either 3 or 10  $\mu$ l of protein solution into two wells A1 and A2.
2. Pipette 0.3 ml of ADV02 into wells A1, A2, B1 and B2.
3. Incubate at room temperature for one minute.
4. Place in reader and measure absorbance at 600 nm (using 580 and 620 nm is possible but the sensitivity will decrease approximately two fold when approaching these limits).
5. Calculate protein concentration based on  $1.00 \text{ OD}_{600 \text{ nm}} = 125 \mu\text{g protein per ml reagent}$ , and multiply this concentration by 100 (for 3  $\mu$ l) or 30 (for 10  $\mu$ l) to achieve  $\mu\text{g protein per ml of original sample}$ . Remember to not use ODs above 1.00 which will be off the standard curve. transfer, block the membrane in TBST/5% non-fat milk (10 mM Tris HCl pH 8.0, 150 mM NaCl, 0.01% Tween 20/5% non-fat milk) for 30 minutes at room temperature.

## VI: Compatibility Table

Analysis of chemicals used in protein biochemistry that may interfere with ADV02 based assays.

<b>Chemical group /chemical name</b>	<b>Tested concentration that does not alter protein assay color (using 10µl per ml ADV02)</b>	<b>Tested concentration that does not alter protein assay response to protein (using 10µl per ml ADV02)</b>
Buffers / Tris pH 8.0	>1.0 M	>1.0 M
Buffers / HEPES pH 8.0	>1.0 M	>1.0 M
Buffers / PIPES pH 7.0	>1.0 M	>1.0 M
Buffers / Potassium phosphate pH7.0	>1.0 M	>1.0 M
Buffers / Sodium bicarbonate pH9.5	>1.0 M	>1.0 M
Reducing agents / BME	>1.0 M	>1.0 M
Reducing agents / DTT	>1.0 M	>1.0 M
Reducing agents / monothioglycerol	>1.0 M	>1.0 M
Denaturants / 8 M urea	100%	100%
Denaturants / 5 M guanidine-HCl	100%	100%
Divalent cations / MgCl <sub>2</sub>	>1.0 M	>1.0 M
Divalent cations / CaCl <sub>2</sub>	>1.0 M	>1.0 M
Divalent cations / NiCl <sub>2</sub>	>0.1 M	>0.1 M
Chelating agents / EDTA	>1.0 M	>1.0 M
Chelating agents / EGTA	>1.0 M	>1.0 M
Detergents / SDS	0.1%	0.1%
Detergents / NP40	10%	10%
Detergents / Triton X-100	10%	0.5%
Detergents / Tween 20	10%	10%
Solvents / DMSO	100%	100%
Solvents / DMF	100%	100%
Solvents / ethyl alcohol	100%	100%
Solvents / methanol	100%	100%
Antifoaming agent / Antifoam-C	10%	10%
Acids / hydrochloric acid	>1.0 M	>0.75 M
Acids / perchloric acid	>1.0 M	>0.75 M
Acids / trichloric acid	>1.0 M	>0.50 M
Acids / nitric acid	>1.0 M	>0.75 M
Acids / sulfuric acid	>1.0 M	>0.25 M

## VII: Troubleshooting

Observation	Possible cause	Solution
<p>1. No increase in blue color</p>	<p>1. Protein concentration too low</p> <p>2. Incorrect labeling of tubes</p>	<p>1. Use the method for dilute or very dilute protein solutions.</p> <p>For protein solutions of less than 0.5 mg/ml it is recommended to use ADV01 which is excellent for detecting low protein concentrations and a good linear response in these low protein concentrations ranges.</p> <p>2. Repeat assay.</p>
<p>2. During measurement of a large number of samples the standards read decreasing absorbance values.</p>	<p>1. Time to read takes too long, i.e. greater than 30 min.</p>	<p>1. Measure fewer samples at a time.</p> <p>2. Or use a high through put method for measurement (e.g. 96-well plates and eight channel pipettors).</p>
<p>3. Buffer blank reads too high (&gt;0.05 OD) absorbance values when compared to ADV02 alone.</p>	<p>1. Buffer contains interfering chemicals (probably detergents).</p>	<p>1. Use the concentrated protein assay.</p> <p>2. Use a detergent free buffer.</p> <p>3. Ethanol precipitate proteins (3 volumes of ethanol), centrifuge (14000 x g 10 min) and resuspend in ADV02.</p>

## VIII: Notes on Updated Version

---

Notes on Updated Version 6.0

A. Manual updated to new manual format.

NOTES:

---

NOTES:

---

