

SARS-CoV-2 Neutralizing Antibody Rapid Test (Colloidal Gold Immunochromatography) Catalog Number 41A260

(Please read this instruction manual before use.)

INTENDED USE

This kit can be used for the qualitative measurement of SARS-CoV-2 neutralizing antibodies in human serum, plasma and whole blood.

SUMMARY

The SARS-CoV-2 virus is a new type of coronavirus of the genus β , encoding four major structural proteins: spike (S), envelope (E), membrane (M) and nucleocapsid (N). The neutralizing antibodies can target the S1 subunit of S protein and bind to the S1 receptor binding domain (RBD), thereby blocking the interaction between S1RBD of the virus and angiotensin-converting enzyme 2 (ACE2) on the surface of human cells, alleviating the threat from SARS-CoV-2. The detection of neutralizing antibodies can be used to monitor the immune response of vaccinated people or people infected by SARS-CoV-2.

ASSAY PRINCIPLE

This assay is based on the specific antibody-antigen reaction with colloidal gold immunochromatography. Colloidal gold particles conjugated with recombinant SARS-CoV-2 S1RBD is pre-coated on the conjugation pad together with colloidal gold conjugated goat anti-chicken IgY as the control particles. The test line (T) on the strip is coated with recombinant SARS-CoV-2 S1RBD, and the control line is coated with chicken IgY. During testing, neutralizing antibodies present in the samples bind to the S1RBD conjugated colloidal gold particles. The complex migrates upward by the capillary effect and is captured by the S1RBD immobilized on the membrane forming the test line (T). Meanwhile, the colloidal gold labeled with goat anti-chicken IgY is captured by chicken IgY forming the control line (C). The result is visible within 15-20 minutes.

REAGENTS AND MATERIALS

Components	1 test/kit	25 tests/kit
SARS-CoV-2 neutralizing antibody test strip	1	25
Assay buffer	0.1ml/tube	2 ml/tube
Quantitative pipette	1	25
Instruction for use	1	1

Note: Different sets of packages are available, please contact us for details.

OTHER MATERIALS REQUIRED, BUT NOT PROVIDED

Timer, Lancets and Alcohol pad

PRECAUTIONS

- All reagents are for *in vitro* diagnostics use only.
- All reagents should be equilibrated to room temperature before use.
- Avoid touching nitrocellulose membrane with your fingers.

-Strips are sensitive to temperature and humidity. The reaction temperature should be at 15 °C-30 °C and the humidity should be below 70%.

-Handle all specimens as if they contain infectious agents. Wear protective clothing such as laboratory coats, disposable gloves and eye protection when specimens are tested.

-Use appropriate volume of sample to ensure the accuracy of this test.

-The quality of expired reagents cannot be guaranteed or if reagents are not stored under required conditions as indicated in the manual.

- Do not use the strip if the pouch is damaged or the seal is broken.

-The used reagents should be discarded according to local regulations.

STORAGE

- The kit needs to be stored at 4 °C-30 °C and in a dry environment.

- Avoiding freezing strips and buffer.

- The test strip is stable until the expiry date only if it has not been opened and kept in the sealed aluminum pouch.

- The assay buffer is valid for 30 days after opening the bottle.

- Do not open the sealed pouch until use. Once opened, the strip should be used within 1 hour.

SPECIMEN COLLECTION AND STORAGE

1. For serum/plasma/whole blood

a. Serum preparation

After collection of the whole blood, allow the blood to clot by leaving it undisturbed at room temperature. This usually takes 15–30 minutes. Remove the clot by centrifuging at 1,000–2,000 x g for 10 minutes in a refrigerated centrifuge.

b. Plasma preparation

Collect whole blood into commercially available anticoagulant-treated tubes e.g., EDTA-treated or citrate-treated.

Cells are removed from plasma by centrifugation for 10 minutes at 1,000–2,000 x g using a refrigerated centrifuge. Centrifugation for 15 minutes at 2,000 x g depletes platelets in the plasma sample.

c. Whole blood

Collect whole blood into commercially available anticoagulant-treated tubes e.g., EDTA-treated or citrate-treated.

d. The serum/plasma can be stored at 2-8°C for 7 days if it cannot be tested in time. For long time storage, serum/plasma samples should be stored below -15°C.

e. Whole blood samples should not be frozen and can be stored at 2-8°C for 3 days.

2. For fingertip whole blood

The fingertip whole blood sample should be tested immediately.

TEST PROCEDURE

Preparation of the test:

Equilibrate kit components in unopened packaging to room temperature (15-30 °C) before starting the test.

Assay procedure:

1. Take the strip out of the packaging bag and place it on the table.
2. Add the sample:



ImmunoDiagnostics Limited

Website: www.immunodiagnosics.com.hk; E-mail: info@immunodiagnosics.com.hk;

Tel: (+852) 3502 2780 / (+86) 0755-2663 1330 ; Fax: (+852) 3502 2781

For serum/plasma/whole blood samples:

Transfer 25-30 µL serum or plasma sample to the sample hole (S) on the test strip using a quantitative pipette.

For fingertip whole blood samples:

a. Clean the puncture site in your fingertip with an alcohol pad.

b. Remove the protective cap of the safety lancet and puncture the fingertip.

c. Transfer 25-30 µL fingertip whole blood to the sample hole (S) on the test strip using a quantitative pipette.

Note:


i. When the sample reaches the arrow indicated in the image to the right, the volume taken up will be approximately 25-30 µL.



ii. Adding too much volume of sample will affect the running of the assay.

3. Remove the vial cap and add 1 drop (30 µL) of the assay buffer to the sample hole.
4. Read the test results within 15-20 minutes. Do not read the results after 25 minutes.

INTERPRETATION OF TEST RESULT



- **Positive result:**
The presence of two lines as the control line (C) and the test line (T) in the result window. A positive result indicates that the neutralizing antibodies are detected in the sample.
- **Negative result:**
The presence of a single line as the control line (C) in the result window. A negative result indicates that the neutralizing antibodies are not detected in the sample.
- **Invalid result:**
If the control line (C) is not visible within the result window after performing the test, the result is invalid. In this case, you should read the instructions carefully and retest the sample.
Note: The test line could be faint. Any line, even if faint, should be interpreted as a line. Do not compare the color intensity of each line to another.

LIMITATIONS OF TEST

- This kit is a qualitative test, and the test results are only for clinical reference, which should not be used as the only basis for clinical diagnosis and treatment.
- Any error in sample collection, processing, transportation and storage may lead to the inaccuracy of the test results.
- Hemolytic samples may cause false positive result. Please avoid using hemolytic samples for this test.

ASSAY PERFORMANCE

Sensitivity and specificity:

The SARS-CoV-2 Neutralizing Antibody Rapid Test has been evaluated with samples obtained from 89 vaccinated

individuals at 14-16 days after their second injection and 220 non-vaccinated subjects. The comparator method Plaque Reduction Neutralization Test (PRNT) using virus (strain HCoV-19/USA/WAI/2020; BEI Resources, Manassas, VA) was used in this trial.

SARS-CoV-2 Neutralizing Antibody Rapid Test (Colloidal Gold Immunochemistry)	PRNT ₅₀		
	Positive	Negative	Total
Positive	86	4	90
Negative	3	216	219
Total	89	220	309
Sensitivity: 86/89	96.63%	(95% CI: 91.13%-99.05%)	
Specificity: 214/220	98.18%	(95% CI: 95.54%-99.47%)	

Limit of detection:

The detection limit was 15 IU/ml, determined by WHO International Standard for anti-SARS-CoV-2 immunoglobulin, NIBSC code:20/136).

Cross-reactivity:

No cross-reactivity was noted when samples of the diseases in the table presented below were tested

Conditions	Conditions
Anti-influenza A IgG	Anti-HBV IgM
Anti-influenza A IgM	ANA
Anti-influenza B IgG	Anti-HIV
Anti-influenza B IgM	Anti-respiratory syncytial virus IgG
Anti-Haemophilus influenzae IgG	Anti-respiratory syncytial virus IgM
Anti-Haemophilus influenzae IgM	Anti-229E
Anti-HCV IgG	Anti-NL63
Anti-HCV IgM	Anti-OC43
Anti-HBV IgG	Anti-HKU1

Interfering substances:

No interference was noted when the substances listed in the table below were tested at the indicated concentration.

Substance	Concentration
Triglyceride	40 mmol/L
Hemoglobin	2 g/L
Bilirubin	350 µmol/L

SYMBOLS

	Manufacturer		EC Declaration of Conformity
	Expiry date		Consult Instruction
	Lot number		Store
	Catalog number		Caution
	In Vitro Diagnostic Device		Name and Address of EU REP

ImmunoDiagnostics Limited
 5F, Biotech Centre 2 (11W), No. 11 Science Park
 West Avenue, Hong Kong Science Park, Shatin. NT, Hong Kong
 E-mail: info@immunodiagnostics.com.hk
 Tel: (+852) 3502 2780; Fax: (+852)3502 2781

