

DATASHEET

Abbexa Ltd, Innovation Centre, Cambridge Science Park, Cambridge, CB4 0EY, UK Telephone: +44 (0) 1223 755950 - Fax: +44 (0) 1223 755951 - E-Mail: info@abbexa.com

Human Indoleamine-2,3-Dioxygenase (IDO1) ELISA Kit

Catalogue No.:abx151900



Human IDO1 ELISA Kit is a sandwich ELISA kit for use with Serum, plasma, tissue homogenates, cell lysates and other biological fluids. This assay has high sensitivity and excellent specificity for detection of Indoleamine-2,3-Dioxygenase (IDO1) No significant cross-reactivity or interference between Indoleamine-2,3-Dioxygenase (IDO1) and analogues was observed.

Please note that this kit is also available as a CLIA Kit abx492858.

Target: Indoleamine-2,3-Dioxygenase (IDO)

Reactivity: Human

Tested Applications: ELISA

Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user.

Test Range: 0.312 ng/ml - 20 ng/ml

Sensitivity: < 0.118 ng/ml

Validity: The validity for this kit is 6 months.

Storage: Store at 2°C to 8°C upon receipt.

Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within

the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested

that the whole assay is performed by the same user throughout.

Standard Form: Lyophilized

ELISA Detection: Colorimetric

ELISA Type: Sandwich



DATASHEET

Abbexa Ltd, Innovation Centre, Cambridge Science Park, Cambridge, CB4 0EY, UK Telephone: +44 (0) 1223 755950 - Fax: +44 (0) 1223 755951 - E-Mail: info@abbexa.com

ELISA Data: Quantitative

Sample Type: Serum, plasma, tissue homogenates, cell lysates and other biological fluids.

Note: This product is for research use only. The range and sensitivity is subject to change. Please

contact us for the latest product information. For accurate results, sample concentrations must be diluted to mid-range of the kit.If you require a specific range, please contact us in advance or write

your request in your order comments.