

Abbexa Ltd, Innovation Centre, Cambridge Science Park, Cambridge, CB4 0EY, UK Telephone: +44 (0) 1223 755950 - Fax: +44 (0) 1223 755951 - E-Mail: info@abbexa.com

DATASHEET

Rabbit Fibrinogen Degradation Product (FDP) ELISA Kit

Catalogue No.:abx155057



Rabbit FDP ELISA Kit is a sandwich ELISA kit for use with Plasma. This assay has high sensitivity and excellent specificity for detection of Fibrinogen Degradation Product (FDP) No significant cross-reactivity or interference between Fibrinogen Degradation Product (FDP) and analogues was observed.

Please note that this kit is also available as a CLIA Kit <u>abx491598</u>.

Target:	Fibrinogen Degradation Product (FDP)
Reactivity:	Rabbit
Tested Applications:	ELISA
Recommended dilutions	: Optimal dilutions/concentrations should be determined by the end user.
Test Range:	0.156 ng/ml - 10 ng/ml
Sensitivity:	< 0.075 ng/ml
Validity:	The validity for this kit is 6 months.
Storage:	Store at 2°C to 8°C upon receipt.
Stability:	The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout.
Standard Form:	Lyophilized
ELISA Detection:	Colorimetric
ELISA Type:	Sandwich

abbexa 🍊

DATASHEET

Abbexa Ltd, Innovation Centre, Cambridge Science Park, Cambridge, CB4 0EY, UK Telephone: +44 (0) 1223 755950 - Fax: +44 (0) 1223 755951 - E-Mail: info@abbexa.com

ELISA Data:	Quantitative
Sample Type:	Plasma.
Note:	This product is for research use only. The range and sensitivity is subject to change. Please contact us for the latest product information. For accurate results, sample concentrations must be diluted to mid-range of the kit. If you require a specific range, please contact us in advance or write your request in your order comments. Please note that our ELISA and CLIA kits are optimised for detection of native samples, rather than recombinant proteins. We are unable to guarantee detection of recombinant proteins, as they may have different sequences or tertiary structures to the native protein.