

Abbexa Ltd, Innovation Centre, Cambridge Science Park, Cambridge, CB4 0EY, UK
Telephone: +44 (0) 1223 755950 - Fax: +44 (0) 1223 755951 - E-Mail: info@abbexa.com

Rat EMMPRIN / CD147 (BSG) ELISA Kit

Catalogue No.: abx155469



Rat EMMPRIN ELISA Kit is a sandwich ELISA kit for use with Serum, plasma, lung lavage fluid and other biological fluids. This assay has high sensitivity and excellent specificity for detection of Extracellular Matrix Metalloproteinase Inducer (EMMPRIIN). No significant cross-reactivity or interference between Extracellular Matrix Metalloproteinase Inducer (EMMPRIIN) and analogues was observed.

Please note that this kit is also available as a CLIA Kit [abx492853](#).

Target: EMMPRIN

Reactivity: Rat

Tested Applications: ELISA

Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user.

Test Range: 15.62 pg/ml - 1000 pg/ml

Sensitivity: < 5.5 pg/ml

Validity: The validity for this kit is 6 months.

Storage: Store at 2°C to 8°C upon receipt.

Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout.

Standard Form: Lyophilized

ELISA Detection: Colorimetric

Abbexa Ltd, Innovation Centre, Cambridge Science Park, Cambridge, CB4 0EY, UK
Telephone: +44 (0) 1223 755950 - Fax: +44 (0) 1223 755951 - E-Mail: info@abbexa.com

ELISA Type: Sandwich

ELISA Data: Quantitative

Sample Type: Serum, plasma, lung lavage fluid and other biological fluids.

Note: This product is for research use only. The range and sensitivity is subject to change. Please contact us for the latest product information. For accurate results, sample concentrations must be diluted to mid-range of the kit. If you require a specific range, please contact us in advance or write your request in your order comments.