



Human eosinophil cationic protein,ECP ELISA Kit

Product Code	CSB-E11729h
Abbreviation	ECP
Target Name	eosinophil cationic protein,ECP
Uniprot No.	P12724
Product Type	ELISA Kit
Immunogen Species	Homo sapiens (Human)
Sample Types	serum, plasma, tissue homogenates
Detection Range	1.56 ng/mL-100 ng/mL
Sensitivity	0.39 ng/mL
Assay Time	1-5h
Sample Volume	50-100ul
Detection Wavelength	450 nm
Lead Time	3-5 working days after you place the order, and it takes another 3-5 days for delivery via DHL or FedEx.
Research Area	Others
Quality Control	<p>A microplate reader capable of measuring absorbance at 450 nm, with the correction wavelength set at 540 nm or 570 nm.</p> <p>An incubator that can provide stable incubation conditions up to 37°C±5°C.</p> <p>Centrifuge</p> <p>Vortex</p> <p>Squirt bottle, manifold dispenser, or automated microplate washer</p> <p>Absorbent paper for blotting the microtiter plate</p> <p>50-300ul multi-channel micropipette</p> <p>Pipette tips</p> <p>Single-channel micropipette with different ranges</p> <p>100ml and 500ml graduated cylinders</p> <p>Deionized or distilled water</p> <p>Timer</p> <p>Test tubes for dilution</p>
Gene Names	RNASE3
Tag Info	quantitative
Protein Description	Sandwich
Component	<p>A micro ELISA plate --- The 96-well plate has been pre-coated with an anti-human ECP antibody. This dismantlable microplate can be divided into 12 x 8 strip plates.</p> <p>Two vials lyophilized standard --- Dilute a bottle of the standard at dilution series, read the OD values, and then draw a standard curve.</p> <p>One vial Biotin-labeled ECP antibody (100 x concentrate) (120 µl/bottle) ---Act</p>



as the detection antibody.

One vial HRP-avidin (100 x concentrate) (120 µl/bottle) --- Bind to the detection antibody and react with the TMB substrate to make the solution chromogenic.

One vial Biotin-antibody Diluent (15 ml/bottle) ---Dilute the high concentration Biotin-antibody to an appropriate working solution.

One vial HRP-avidin Diluent (15 ml/bottle) ---Dilute the high concentration HRP-avidin solution to an appropriate solution.

One vial Sample Diluent (50 ml/bottle)---Dilute the sample to an appropriate concentration.

One vial Wash Buffer (25 x concentrate) (20 ml/bottle) --- Wash away unbound or free substances.

One vial TMB Substrate (10 ml/bottle) --- Act as the chromogenic agent. TMB interacts with HRP, eliciting the solution turns blue.

One vial Stop Solution (10 ml/bottle) --- Stop the color reaction. The solution color immediately turns from blue to yellow.

Four Adhesive Strips (For 96 wells) --- Cover the microplate when incubation. An instruction manual

Description

This Human ECP ELISA Kit was designed for the quantitative measurement of Human ECP protein in serum, plasma, tissue homogenates. It is a Sandwich ELISA kit, its detection range is 1.56 ng/mL-100 ng/mL and the sensitivity is 0.39 ng/mL.

Product Precision

Intra-assay Precision (Precision within an assay): CV%<8%

Three samples of known concentration were tested twenty times on one plate to assess.

Inter-assay Precision (Precision between assays): CV%<10%

Three samples of known concentration were tested in twenty assays to assess.

Linearity

To assess the linearity of the assay, samples were spiked with high concentrations of human ECP in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.

?	Sample	Serum(n=4)
1:1	Average %	93
	Range %	88-98
1:2	Average %	97
	Range %	92-103
1:4	Average %	100
	Range %	94-108
1:8	Average %	108
	Range %	103-112

Recovery


The recovery of human ECP spiked to levels throughout the range of the assay in various matrices was evaluated. Samples were diluted prior to assay as directed in the Sample Preparation section.

Sample Type	Average % Recovery	Range
Serum (n=5)	90	87-95
EDTA plasma (n=4)	92	89-95



Typical

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



ng/ml	OD1	OD2	Average	Corrected
100	2.500	2.493	2.497	2.392
50	2.003	1.998	2.001	1.896
25	1.275	1.266	1.271	1.166
12.5	0.587	0.573	0.580	0.475
6.25	0.369	0.357	0.363	0.258
3.12	0.221	0.208	0.215	0.110
1.56	0.176	0.169	0.173	0.068
0	0.108	0.102	0.105	?

Msd

{\"0\":{\"fileurl\":\"https://www.cusabio.com/uploadfile/msds/MSDS CSB-E11729h.pdf\", \"filename\":\"MSDS\"}}